

Implantable cardioverter defibrillator therapy for primary prevention of sudden cardiac death in the real world: Main findings from the French multicentre DAI-PP programme (pilot phase)

Abbreviated title: ICDs for primary prevention of sudden cardiac death

Serge Boveda^a, Rodrigue Garcia^{b,c}, Pascal Defaye^d, Olivier Piot^e, Kumar Narayanan^{f,g}, Sergio Barra^{h,i}, Daniel Gras^j, Rui Providencia^{a,k}, Vincent Algalarrondo^l, Frankie Beganton^f, Marie-Cécile Perier^f, Sophie Jacob^m, Pierre Bordacharⁿ, Dominique Babuty^o, Didier Klug^p, Christophe Leclercq^q, Laurent Fauchier^o, Nicolas Sadoul^r, Jean-Claude Deharo^s, Eloi Marijon^{f,t,u,*}, for the DAI-PP investigators

^a *Clinique Pasteur, 31076 Toulouse, France*

^b *CHU Poitiers, 86021 Poitiers, France*

^c *Université de Poitiers, 86073 Poitiers, France*

^d *CHU Michallon, 38700 La Tronche, France*

^e *Centre Cardiologique du Nord, 93200 Saint-Denis, France*

^f *Paris Cardiovascular Research Centre (Inserm U970), 75015 Paris, France*

^g *Maxcure Hospitals, Hyderabad, Telangana 500081, India*

^h *Royal Papworth Hospital NHS Foundation Trust, Cambridge University Health Partners, Cambridge CB2 0AY, UK*

ⁱ *Hospital da Luz Arrabida, 4400-346 Vila Nova de Gaia, Portugal*

^j *Nouvelles Cliniques Nantaises, 44200 Rezé, France*

^k *Barts Heart Centre, Barts Health NHS Trust, London EC1A 7BE, UK*

^l *Hôpital Bichat-Claude-Bernard, 75018 Paris, France*

^m *IRSN, 92260 Fontenay aux Roses, France*

ⁿ *CHU Haut Lévêque, 33600 Pessac, France*

^o *CHU Trousseau, 37170 Chambray-lès-Tours, France*

^p *CHRU Lille, 59000 Lille, France*

^q *CHU Pontchaillou, 35033 Rennes, France*

^r *CHU Brabois, 54500 Vandœuvre-lès-Nancy, France*

^s *Hôpital de La Timone, 13005 Marseille, France*

^t *European Georges Pompidou Hospital, 75908 Paris, France*

^u *Paris Descartes University, 75006 Paris, France*

* Corresponding author at: Unité de Rythmologie, Département de Cardiologie, Hôpital Européen Georges Pompidou, 20–40 Rue Leblanc, 75908 Paris Cedex 15, France.

E-mail address: eloi_marijon@yahoo.fr (E. Marijon). Fax: +33 1 5609 3047

Serge Boveda and Rodrigue Garcia are joint first authors.

Summary

This review summarizes the main findings of the French multicentre DAI-PP pilot programme, and discusses the related clinical and research perspectives. This project included retrospectively more than 5000 subjects with structural heart disease who received an implantable cardioverter defibrillator (ICD) for primary prevention of sudden cardiac death, and followed for a mean period of 3 years. This collaborative research programme led to 20 publications. Several useful and clinically relevant findings emerged from this cohort, which are summarized in this review. Firstly, our registry revealed the changing trends in primary prevention ICD implantation over the past decade, with increases in age at implantation and the proportions of cardiac resynchronization therapy, and female recipients. The occurrence of early complications after implantation remained significantly associated with global mortality (odds ratio 1.70; $P = 0.003$). Female recipients of resynchronization therapy had a lower incidence of appropriate therapies (hazard ratio 0.62; $P < 0.001$ versus males). Overall mortality was higher among patients with ischaemic compared with non-ischaemic cardiomyopathy (hazard ratio 1.31; $P = 0.01$); incidences of appropriate therapies were similar. During follow-up, 3.2% of ICD recipients had a heart transplantation. Looking at utility of ICD implanted after 80 years, 19.4% received at least one appropriate therapy before death. Compared with single-chamber ICD, dual-chamber ICD did not reduce rates of inappropriate therapies, and were associated with a higher rate of complications. Finally, this pilot phase have led to a prospective evaluation since May 2018, assessing ICD therapy in the primary and secondary prevention, in patients with structural and electrical heart diseases followed-up with a remote monitoring dedicated platform.

Résumé

Cet article résume les principaux résultats de la phase pilote du programme DAI-PP et discute ses perspectives scientifiques. Ce projet a inclus plus de 5000 sujets ayant une cardiopathie et implantés d'un défibrillateur automatique implantable (DAI) en prévention primaire de la mort subite avec un suivi moyen de 3 ans. Ce programme de recherche collaboratif a abouti à 20 publications dont les principaux résultats sont résumés dans cette revue. Premièrement, ce registre a évalué l'évolution des implantations de DAI sur la dernière décennie, décrivant une augmentation de l'âge à l'implantation, de la proportion de resynchronisation cardiaque, et de femmes implantées. La survenue d'une complication précoce était significativement associée à la mortalité globale (OR

1,70 ; $P = 0,003$). L'incidence des thérapies appropriées était plus faible chez les femmes implantées d'une resynchronisation (HR 0,62 ; $P < 0,001$ versus hommes). La mortalité globale était plus élevée chez les porteurs d'une cardiopathie ischémique comparé aux non-ischémiques alors que l'incidence des thérapies appropriées était similaire. Durant le suivi, 3,2 % des patients ont bénéficié d'une transplantation cardiaque. Chez les patients ≥ 80 ans à l'implantation, 19,4 % ont reçu au moins une thérapie appropriée avant leur décès. Enfin, comparé aux appareils mono-chambre, les DAI double-chambre ne réduisaient pas le taux de thérapies inappropriées et étaient associés avec un risque plus élevé de complications. Après cette évaluation pilote, un programme prospectif a été initié en France (mai 2018) visant à évaluer les DAI en prévention primaire et secondaire, chez des patients porteurs de cardiopathie structurelle et électrique, suivis en télécardiologie via une plateforme dédiée.

KEYWORDS

Implantable cardioverter defibrillator;

Primary prevention;

Sudden cardiac death;

Complications;

Epidemiology

MOTS CLÉS :

Défibrillateur automatique implantable ;

Prévention primaire ;

Mort subite cardiaque ;

Complications ;

Épidémiologie

Abbreviations: CI, confidence interval; CRT, cardiac resynchronization therapy; HR, hazard ratio; ICD, implantable cardioverter defibrillator; ICM, ischaemic cardiomyopathy; NICM, non-ischaemic cardiomyopathy; NYHA, New York Heart Association; OR, odds ratio.

Background

Sudden cardiac death remains a major mode of death in Western world [Ref sup 1 and 2]. Implantable cardioverter defibrillators (ICDs) for primary prevention of sudden cardiac death have been demonstrated to be associated with a significant reduction in overall mortality, among selected patients with dilated and ischaemic cardiomyopathies [1, 2]. Nevertheless, the implantation of such devices has the potential for numerous complications [3, 4]. Moreover, accurate prediction of sudden cardiac death is still a considerable challenge, and the number of patients implanted with a primary prevention ICD who will never experience ventricular arrhythmia represents both wasted effort and unnecessary risk, underlining the need for continuous improvements in the selection process for ICD candidates [5].

Although randomized controlled trials have given us a wealth of evidence, long-term data from the real-world setting are valuable and complementary. With this in mind, the pilot phase of the DAI-PP programme was conceived as a retrospective French multicentre registry aiming to assess the characteristics and outcomes of all patients implanted with an ICD for the primary prevention of sudden cardiac death between 2002 and 2012 (ClinicalTrials.gov Identifier: NCT#01992458), across 12 academic and private electrophysiology centres. All data were recorded retrospectively according to a specific protocol. Overall, 5539 consecutive patients were included, and the mean follow-up was 3 years. This registry led to 20 publications, and gave insights into several practical aspects of ICD use in primary prevention. This paper summarizes the main findings from the different DAI-PP analyses, and puts these into perspective, to help to eventually optimize primary prevention of sudden cardiac death using ICD therapy.

Main findings (Table 1)

Trends in primary prevention ICD implantation over a decade

As technology and clinical practice surrounding the use of primary prevention ICD implantation are in a state of constant evolution, the temporal trends in characteristics and outcomes of ICD therapy were assessed between 2002 and 2012 [6]. In addition to a shift in the type of device implanted, significant

increases were observed in mean age (from 61.5 ± 11.6 to 63.2 ± 10.9 years; $P = 0.0016$) and the proportions of cardiac resynchronization therapy (CRT) with an ICD (from 43.6% to 60.4%; $P = 0.0001$), non-ischaemic cardiomyopathy (NICM) (from 31.0% to 44.7%; $P < 0.0001$) and female recipients (from 11.4% to 15.8%; $P = 0.004$). Reductions in annual mortality incidence (from 5.4% to 4.3%; $P = 0.05$) and appropriate therapy incidence (from 10.4% to 7.1%; $P = 0.0004$) were seen over the decade. By contrast, the incidence of ICD-related late complications (> 30 days after implantation) increased significantly (from 4.6% to 7.6%; $P = 0.003$). This analysis highlighted significant changes in the patterns of use and outcomes in primary prevention ICD implantation, with reductions in mortality and appropriate therapies, counterbalanced by an increase in complications. While the former is likely to be a reflection of advances in and better adherence to guideline-directed medical therapies, the latter is concerning, and may be the result of increased numbers of implantations in older patients with more co-morbidities. Of note, in this large cohort of primary prevention ICD recipients, more than a quarter were implanted without fulfilling proper guideline criteria, similar to rates reported in other large registries [Ref sup 3]. Despite a less severe clinical profile at implantation and lower mortality rates, off-guideline patients had, rather surprisingly, similar rates of appropriate therapies. These patients were also likely to benefit from their ICDs, underlying the need for a better risk stratification in the future.

Special focus on specific groups: Women, advanced functional heart failure class and the elderly

Through the DAI-PP project we were able to assess outcomes according to specific variables, such as sex, New York Heart Association (NYHA) class and age. In most ICD trials, women were markedly under-represented. In our registry, women constituted a minority of the recipients of an ICD as well (15.1%), 53.8% of whom received CRT [7]. Compared with men, women presented a different clinical profile, with a significantly higher proportion of NICM (60.2% vs 36.2%; $P < 0.001$), a wider QRS complex (QRS > 120 ms: 74.6% vs 68.5%; $P = 0.003$) a higher NYHA functional class and a lower prevalence of atrial fibrillation. During follow-up, female recipients of an ICD had a lower incidence of appropriate ICD therapies (hazard ratio [HR] 0.59, 95% confidence interval [CI] 0.45–0.76; $P < 0.001$) and all-cause death (HR 0.87, 95% CI 0.66–1.15; $P = 0.324$) than their male counterparts, whereas the rates of inappropriate

shocks and early complications were similar in both groups. Scientific data on sex-specific outcomes is still growing, and pooled data might allow to get enough power, to eventually lead to sex-specific ICD guidelines in the future [[8]Ref sup 4].

Although guidelines use NYHA class in the decision process for ICD implantation, mortality benefit in NYHA class III patients (especially compared to NYHA class II) is a matter of debate [1, 2]. Analysis of outcomes by NYHA class in the DAI-PP database demonstrated that although patients in NYHA class III have higher overall mortality (HR per NYHA class 1.63, 95% CI 1.11–2.41; $P = 0.014$), driven by an increase in cardiovascular death, they experienced a similar incidence of appropriate ICD therapies [9]. Incidence of ICD-unresponsive sudden death remained very low, and was also similar across NYHA classes, supporting the efficacy of ICD in relatively advanced heart failure.

The benefit of primary prevention ICD implantation is also controversial in elderly patients, because of the greater competing risk of non-arrhythmic death. For this purpose, patients' characteristics and outcomes were analysed in three age groups: 18–59 years; 60–74 years; and ≥ 75 years [10]. As expected, older patients exhibited higher global mortality after ICD implantation. However, rates of sudden death and appropriate device therapies were similar across age groups. Older age was independently associated with a higher rate of early postimplantation complications (odds ratio [OR] 1.28 for age 60–74 years; OR 1.49 for age ≥ 75 years; $P = 0.03$) and a lower rate of inappropriate therapies. Among patients aged ≥ 80 years, 19.4% received at least one appropriate therapy before death [11]. Our results might suggest that among well-selected older subjects with relatively few co-morbidities, as reflected in our study population, primary prevention ICD implantation may be of relevance. Nevertheless, the inherent limitations in interpreting observational data in this particular competing-risk situation call for randomized controlled trials to provide definitive answers. Meanwhile, a careful multidisciplinary evaluation is needed to guide patient selection for ICD implantation in the elderly population.

Significant impact of early periprocedural complications

To date, the extent to which early complications after ICD implantation are associated with morbidity and mortality at longer term, in daily practice, has only been addressed in limited populations. The DAI-PP project assessed the incidence and consequences of early complications (≤ 30 days) after ICD

implantation [12]. Overall, early complications occurred in 13.5% of patients, with the most frequent complications being bleeding related (hematoma, 35.9%) and lead dislodgment (20.7%). Independent factors associated with occurrence of early complications were severe renal impairment (OR 1.66; $P = 0.02$), age ≥ 75 years (OR 1.01; $P = 0.03$), CRT (OR 1.58; $P = 0.01$) and anticoagulant therapy (OR 1.28; $P = 0.03$). After adjustment, early complications remained associated with mortality (OR 1.70; $P = 0.003$) (Fig. 1). These results reinforce the impact of early complications, and particularly avoidance of pocket haematoma, which was the most frequent early complication. While recent guidelines demonstrated a high level of evidence regarding vitamin K antagonist management, scientific data is still limited regarding non-vitamin K antagonist oral anticoagulant management in the periprocedural period of implantable cardiac electronic devices [13].

Perioperative defibrillation testing

Recent literature suggests that routine defibrillation testing does not improve defibrillation efficacy or reduce arrhythmic death [14]. Nevertheless, the impact of defibrillation testing had not been specifically evaluated in recipients of CRT with a ICD implanted for primary prevention of sudden cardiac death. Therefore, we compared outcomes between patients who underwent defibrillation testing and those who did not, immediately after the implantation of a CRT ICD [15]. Out of the 1516 patients, defibrillation testing was performed in 958 (63%) patients. All of the three perioperative deaths occurred in the group that had defibrillation testing, and were related to defibrillation testing itself. The adjusted incidence of overall mortality was lower in the group that had defibrillation testing (HR 0.6; $P < 0.0001$), mainly reflecting the less severe cardiac disease in this group (making patients more likely to be selected for defibrillation testing), rather than a lower rate of ICD-unresponsive sudden death during follow-up. Taken overall, our results do not encourage systematic defibrillation testing in recipients of CRT with a ICD.

Outcomes comparison between NICM and ischaemic cardiomyopathy (ICM)

The benefit of primary prevention ICD implantation in NICM has been debated recently, and the evidence level for ICD implantation in NICM is considered to be lower than for ICM in the European Society of Cardiology guidelines [16, 17]. Consequently, the DAI-PP study group compared outcomes between recipients of an ICD implanted for primary prevention with NICM (40% of patients) and with ICM (60% of patients) [18]. Overall mortality was higher in the ICM group than the NICM group (adjusted HR 1.31; $P = 0.01$), and the increase in cardiac mortality in the ICM group was mainly the result of non-cardiovascular causes ($P = 0.0002$). The incidence of appropriate ICD therapy was similar in the two groups. These observational data appear to support the efficacy of ICD in NICM, similar to other recent registry data and meta-analyses [ref sup 8].

CRT: Is an additional ICD always required?

The DAI-PP project also collaborated with a large multicentre European cohort, assessing the added value of an ICD over CRT in ICM and NICM [19]. Overall, 5307 consecutive patients with NICM or ICM underwent CRT implantation with ($n = 4037$) or without ($n = 1270$) a ICD. After a mean follow-up period of 41.4 ± 29.0 months, patients with ICM had better survival when receiving CRT with a ICD compared with those who received CRT without a ICD (HR 0.76; $P = 0.005$), whereas in patients with NICM, no such difference was observed (HR 0.92; $P = 0.49$) (Fig. 2). Outcome differences according to sex category were also assessed, and benefit from CRT with a ICD was seen in male patients (HR 0.78; $P = 0.012$), but not in female patients (HR 0.87; $P = 0.43$). These findings were confirmed in another collaborative work assessing sex-specific outcomes with the addition of a ICD to CRT [20]. These results reinforce the importance of careful patient selection, especially among women, to optimize the benefit and cost-effectiveness of an added ICD in patients with an indication for CRT.

In addition further analyses were carried out from a large European CRT consortium comprising French, UK, Czech and Swedish patients who received CRT implantation or upgrade between 2002 and 2012, in order to provide long term outcomes, especially with regard to the occurrence of sudden cardiac death, in CRT patients without and with defibrillator. Progressive heart failure death was found to represent the most frequent cause of death in patients

surviving the first five years after CRT implant. By contrast, SCD represented a very low proportion of late mortality irrespective of the presence of an ICD.

[Ref sup 5]

Reducing the ICD shock burden

The benefits of ICD implantation in the primary prevention of sudden cardiac death are offset to some extent because of the morbidity conferred by both appropriate and, more importantly, inappropriate shocks. The MADIT-RIT trial suggested very-high-rate cut-off and long detection time strategies to reduce inappropriate shocks [21]. We assessed the real-world efficacy of a very-high-rate cut-off programming strategy in our registry, by comparing 500 patients programmed with a very-high-rate cut-off (VH-RATE group: monitor zone 170–219 beats/min; ventricular fibrillation zone ≥ 220 beats/min with 13 ± 4 detection intervals) with 1500 matched control patients programmed with standard one therapy or two therapy zones [22]. VH-RATE programming was associated with reductions in appropriate therapy (HR 0.40; $P < 0.0001$) and inappropriate shock risk (HR 0.42; $P < 0.0001$) (Fig. 3). Nevertheless, there was no significant difference in overall survival between the groups. Optimal ICD programming, which focuses on the reduction of therapies, could be a safe and effective strategy in preventing subsequent morbidity and/or mortality from ICD shocks.

Efficacy of single-chamber versus dual-chamber ICDs

Another putative way to reduce inappropriate shocks could be the implantation of an atrial lead, to enhance reliability of supraventricular arrhythmia discrimination. However, the benefit of dual-chamber versus single-chamber ICDs could be counterbalanced by a higher rate of complications related to dual-chamber ICD implantation. Using the DAI-PP dataset, we compared 1258 recipients of single-chamber ICDs with 1280 recipients of dual-chamber ICDs [23]. While the rates of periprocedural complications were higher in the dual-chamber ICD group (12.1%) compared with the single-chamber ICD group (8.8%; $P =$

0.008), the proportions of appropriate therapies (24.7% vs 23.8%), inappropriate shocks (8.4% vs 7.8%) and all-cause mortality (12.4% vs 13.2%) were similar between groups (Fig. 4). Thus, our DAI-PP experience does not support the routine implantation of dual-chamber ICDs to reduce inappropriate therapies.

Improving the benefit of ICD therapy by improving selection accuracy

Better patient selection, particularly with regard to identifying those at high risk of non-arrhythmic death, may improve the cost-effectiveness of primary prevention ICD implantation [24]. In this context, a cause-of-death analysis was carried out in 2485 patients, to investigate the extent to which a prognostic score derived from a randomized controlled trial in patients with coronary artery disease was of potential interest in the real-world setting for patients with ICM and NICM [25]. The risk score included points for NYHA functional class III or greater, age > 70 years, QRS duration > 120 ms, atrial fibrillation and glomerular filtration rate < 60 mL/min. After a mean follow-up of 3.0 years, the overall mortality rate was 5.9 per 100 patient-years, which increased progressively with the number of risk factors. The higher mortality rate among patients with the highest scores resulted from an increase in non-arrhythmic mortality (from 2.1 to 14.8 per 100 patient-years; $P < 0.001$), whereas the occurrence of appropriate ICD therapies did not change significantly across the categories. The C-statistic testing of the score was observed to be highly similar for patients with ICM (0.685) and NICM (0.658) and those receiving CRT (0.678). These findings underline the feasibility in the real-world setting of identifying patients who are at significant risk of non-arrhythmic death among those eligible for primary prevention ICD implantation. A careful analysis of the risk-benefit ratio and additional discussion with the patient would be warranted in such identified cases.

Electrical storm in primary prevention recipients of an ICD

The characteristics and risk of death in patients who developed electrical storm during follow-up were evaluated [26]. Occurrence of electrical storm was associated with poor prognosis, as the probability of death was higher after electrical storm (HR 3.77; $P < 0.001$). In the multivariable survival analysis, renal failure (HR 1.8; $P = 0.007$), left ventricular ejection fraction < 30% (HR 1.7; $P = 0.03$) and male sex (HR 2.3; $P = 0.05$) were associated with the occurrence of

electrical storm. These data underline the prognostic implications of electrical storm, which may be mitigated nowadays by advances in catheter ablation techniques and by neural modulation, and prevented in some extent with the optimization of pharmacological therapies.

CRT response, survival and heart transplantation in recipients of an ICD

Almost a third of patients with heart failure fail to respond to CRT. To help identify such patients, a simple clinical risk score called ScREEN was developed using the DAI-PP cohort, and then validated in an external cohort [27]. This score used the following variables: sex category; renal function; electrocardiogram/QRS width; ejection fraction; and NYHA class. External validation showed good calibration, accuracy, and discrimination (C-statistic 0.67), with CRT response increasing progressively from 37.5% in patients with a score of 0 to 91.9% in those with a score of 5 ($P < 0.001$). Finally, our real-world data support other findings from other groups. As expected, there were differences in the risk of death and probability of CRT response among comorbidity groups but, the burden of comorbidity did not appear to significantly influence the probability of response to CRT, despite a potentially smaller survival benefit in patients with ≥ 3 comorbidities. Although ICD implantation should be carefully considered in patients with multiple comorbidities, CRT should not be withheld in this context of heart failure patient with enlarged QRS. [Ref sup 6]

While ICDs are not recommended for patients with an expected survival of < 1 year, the factors in real life associated with such an outcome have not been well studied. We assessed the prevalence and factors associated with survival of ≤ 1 year among patients implanted with an ICD [Ref sup 7]: 230 out of 5457 (4.2%) patients survived for ≤ 1 year. Causes of death were similar in those who survived ≤ 1 year versus > 1 year, and patients who survived for ≤ 1 year had fewer appropriate and inappropriate ICD therapies than patients who survived for > 1 year (respectively, 14% vs 23% [$P = 0.004$] and 2% vs 7% [$P = 0.009$]). In the multivariable analysis, older age, higher NYHA class (\geq III), atrial fibrillation and lower left ventricular ejection fraction were significantly associated with survival for ≤ 1 year.

Finally, the frequency, characteristics and outcomes of primary prevention recipients of an ICD who undergo heart transplantation during follow-up were described in the DAI-PP study [28]: 176 (3.2%) patients underwent heart transplantation. The median duration between ICD implantation and heart transplantation was 484 days (interquartile range 169–1117 days). The incidence of appropriate ICD therapies was relatively high in patients who had a heart transplantation compared with those who did not (92.7 vs 76.1 per 1000 person-years; $P = 0.64$), reinforcing the importance of specific cardiac rhythm management for these patients, while they await heart transplantation.

Conclusions

The pilot phase of DAI-PP programme has provided valuable information on several practical and clinically relevant aspects of primary prevention ICD implantation in the real-world population of patients with ischaemic and non-ischaemic cardiomyopathy. This pilot led to a prospective evaluation which started in May 2018, assessing ICD therapy in the primary and secondary prevention, in patients with structural and electrical heart diseases, with a remote monitoring follow-up using a dedicated platform. That should further enhance our understanding of sudden cardiac death, to eventually optimize the field of preventive actions.

Acknowledgements

We thank the investigators of the enrolling medical centres for participating in the study and sharing their data, the Research Associates Nicolas Estrugo, Sandrine Hervouet, Nathalie de Carsin, Radu Mosei, Juliette Tennenbaum, Raoul Hubac-Coupet, Alexandre Bendavid and Marine Sroussi for collecting the data, and the Arrhythmia Group of the French Society of Cardiology for supporting this pilot study.

Sources of funding

This work was supported by the following independent institutions: the French Institute of Health and Medical Research; the Toulouse Association for the Study of Rhythm Disturbances; and the French Society of Cardiology.

Disclosure of interest

S. B. Consultancy fees from the companies **Medtronic, Boston Scientific** and **Microport**.

R. G. Consultancy fees from the companies **Boston Scientific** and **St. Jude Medical**.

P. D. Grants and honoraria from the companies **Boston Scientific, Medtronic, Abbott** and **Microport**.

O. P. Consultancy honoraria from the companies **Medtronic** and **St. Jude Medical**. Research grants from the companies **Medtronic** and **Boston Scientific**.

D. G. Consultancy fees from the companies **Abbott, Biotronik, Boston Scientific** and **Medtronic**.

D. B. Travel support and clinical study support from the companies **Biotronik, Boston Scientific, Medtronic, St. Jude Medical** and **Microport**.

D. K. Consultancy fees from the companies **St. Jude Medical, Medtronic, Microport, Boston Scientific** and **Biotronik**.

L. F. Consultant or speaker for the companies **Bayer, BMS/Pfizer, Boehringer Ingelheim, Medtronic, Novartis**.

N. S. Personal fees from the companies **Biotronik, Boston Scientific, Medtronic, St. Jude Medical** and **Microport**.

E. M. Research grants, consultancy fees and honoraria from the companies **Zoll, Boston Scientific, Abbott, Microport, Medtronic, Boston Scientific, Bristol-Myers Squibb/Pfizer, Boehringer Ingelheim, Bayer, AstraZeneca** and **Mylan**.

The other authors declare that they have no conflicts of interest concerning this article.

References

- [1] Bardy GH, Lee KL, Mark DB, et al. Amiodarone or an implantable cardioverter-defibrillator for congestive heart failure. *N Engl J Med* 2005;352:225-37.
- [2] Moss AJ, Zareba W, Hall WJ, et al. Prophylactic implantation of a defibrillator in patients with myocardial infarction and reduced ejection fraction. *N Engl J Med* 2002;346:877-83.
- [3] de Bie MK, van Rees JB, Thijssen J, et al. Cardiac device infections are associated with a significant mortality risk. *Heart Rhythm* 2012;9:494-8.
- [4] Kleemann T, Becker T, Doenges K, et al. Annual rate of transvenous defibrillation lead defects in implantable cardioverter-defibrillators over a period of >10 years. *Circulation* 2007;115:2474-80.
- [5] Betts TR, Sadarmin PP, Tomlinson DR, et al. Absolute risk reduction in total mortality with implantable cardioverter defibrillators: analysis of primary and secondary prevention trial data to aid risk/benefit analysis. *Europace* 2013;15:813-9.
- [6] Boveda S, Narayanan K, Jacob S, et al. Temporal trends over a decade of defibrillator therapy for primary prevention in community practice. *J Cardiovasc Electrophysiol* 2017;28:666-73.
- [7] Providencia R, Marijon E, Lambiase PD, et al. Primary prevention implantable cardioverter defibrillator (ICD) therapy in women – Data from a multicenter French registry. *J Am Heart Assoc* 2016;5:e002756.
- [8] Barra S, Providencia R, Boveda S, et al. Do women benefit equally as men from the primary prevention implantable cardioverter-defibrillator? *Europace* 2018;20:897-901.
- [9] Providencia R, Boveda S, Defaye P, et al. Outcome of primary prevention implantable cardioverter defibrillator therapy according to New York Heart Association Functional classification. *Am J Cardiol* 2016;118:1225-32.
- [10] Fauchier L, Marijon E, Defaye P, et al. Effect of age on survival and causes of death after primary prevention implantable cardioverter-defibrillator implantation. *Am J Cardiol* 2015;115:1415-22.

- [11] Zakine C, Garcia R, Narayanan K, et al. Prophylactic implantable cardioverter-defibrillator in the very elderly. *Europace* 2019;19. doi: 10.1093/europace/euz041.
- [12] Ascoeta MS, Marijon E, Defaye P, et al. Impact of early complications on outcomes in patients with implantable cardioverter-defibrillator for primary prevention. *Heart Rhythm* 2016;13:1045-51.
- [13] Sticherling C, Marin F, Birnie D, et al. Antithrombotic management in patients undergoing electrophysiological procedures: a European Heart Rhythm Association (EHRA) position document endorsed by the ESC Working Group Thrombosis, Heart Rhythm Society (HRS), and Asia Pacific Heart Rhythm Society (APHS). *Europace* 2015;17:1197-214.
- [14] Healey JS, Hohnloser SH, Glikson M, et al. Cardioverter defibrillator implantation without induction of ventricular fibrillation: a single-blind, non-inferiority, randomised controlled trial (SIMPLE). *Lancet* 2015;385:785-91.
- [15] Perrin T, Mechulan A, Boveda S, et al. Does defibrillation testing influence outcomes after CRT-D implantation? A cause-of-death analysis from the DAI-PP study. *Int J Cardiol* 2016;221:951-6.
- [16] Kober L, Thune JJ, Nielsen JC, et al. Defibrillator implantation in patients with nonischemic systolic heart failure. *N Engl J Med* 2016;375:1221-30.
- [17] McMurray JJ, Adamopoulos S, Anker SD, et al. ESC guidelines for the diagnosis and treatment of acute and chronic heart failure 2012: The Task Force for the Diagnosis and Treatment of Acute and Chronic Heart Failure 2012 of the European Society of Cardiology. Developed in collaboration with the Heart Failure Association (HFA) of the ESC. *Eur J Heart Fail* 2012;14:803-69.
- [18] Amara N, Boveda S, Defaye P, et al. Implantable cardioverter-defibrillator therapy among patients with non-ischaemic vs. ischaemic cardiomyopathy for primary prevention of sudden cardiac death. *Europace* 2018;20:65-72.
- [19] Barra S, Boveda S, Providencia R, et al. Adding defibrillation therapy to cardiac resynchronization on the basis of the myocardial substrate. *J Am Coll Cardiol* 2017;69:1669-78.

- [20] Barra S, Providencia R, Duehmke R, et al. Sex-specific outcomes with addition of defibrillation to resynchronisation therapy in patients with heart failure. *Heart* 2017;103:753-60.
- [21] Sedlacek K, Ruwald AC, Kutiyifa V, et al. The effect of ICD programming on inappropriate and appropriate ICD Therapies in ischemic and nonischemic cardiomyopathy: the MADIT-RIT trial. *J Cardiovasc Electrophysiol* 2015;26:424-33.
- [22] Clementy N, Challal F, Marijon E, et al. Very high rate programming in primary prevention patients with reduced ejection fraction implanted with a defibrillator: Results from a large multicenter controlled study. *Heart Rhythm* 2017;14:211-7.
- [23] Defaye P, Boveda S, Klug D, et al. Dual- vs. single-chamber defibrillators for primary prevention of sudden cardiac death: long-term follow-up of the Defibrillateur Automatique Implantable-Prevention Primaire registry. *Europace* 2017;19:1478-84.
- [24] Goldenberg I, Vyas AK, Hall WJ, et al. Risk stratification for primary implantation of a cardioverter-defibrillator in patients with ischemic left ventricular dysfunction. *J Am Coll Cardiol* 2008;51:288-96.
- [25] Providencia R, Boveda S, Lambiase P, et al. Prediction of nonarrhythmic mortality in primary prevention implantable cardioverter-defibrillator patients with ischemic and non-ischemic cardiomyopathy. *JACC Clin Electrophysiol* 2015;1:29-37.
- [26] Escande W, Marijon E, Defaye P, et al. Electrical storms in patients with implantable cardioverter-defibrillators for primary prevention. *J Am Coll Cardiol* 2016;68:1248-50.
- [27] Providencia R, Marijon E, Barra S, et al. Usefulness of a clinical risk score to predict the response to cardiac resynchronization therapy. *Int J Cardiol* 2018;260:82-7.
- [28] Algalarrondo V, Perault R, Bories MC, et al. Prophylactic implantable cardioverter defibrillators for primary prevention: From implantation to heart transplantation. *Arch Cardiovasc Dis* 2018;111:758-65.

TO ADD:

[Ref sup 1] Waldmann V, Bougouin W, Karam N, et al. Sudden cardiac death: A better understanding for a better prevention. *Ann Cardiol Angeiol (Paris)*. 2017 Sep;66(4):230-238.

[Ref sup 2] Jouven X, Bougouin W, Karam N, Marijon E. Epidemiology of sudden cardiac death: Data from the Paris-Sudden Death Expertise Center registry. *Rev Prat*. 2015 Sep;65(7):916-8.

[Ref sup 3] Escande W, Boveda S, Defaye P, et al. DAI-PP Investigators. Outcomes in guideline-based versus off-guideline primary prevention implantable cardioverter-defibrillator recipients. *J Am Coll Cardiol*. 2017 Sep 5;70(10):1302-1303.

[Ref sup 4] Karam N, Marijon E, Bougouin W, Spaulding C, Jouven X. Sudden cardiac death: Are women different.?. *Ann Cardiol Angeiol (Paris)*. 2016 Dec;65(6):390-394.

[Ref sup 5] Barra S, Duehmke R, Providência R, Narayanan K, Reitan C, Roubicek T, Polasek R, Chow A, Defaye P, Fauchier L, Piot O, Deharo JC, Sadoul N, Klug D, Garcia R, Dockrill S, Virdee M, Pettit S, Agarwal S, Borgquist R, Marijon E, Boveda S. Very long-term survival and late sudden cardiac death in cardiac resynchronization therapy patients. *Eur Heart J*. 2019 May 2. pii: ehz238. doi: 10.1093/eurheartj/ehz238.

[Ref sup 6] Barra S, Providência R, Agarwal S, Marijon E, Boveda S. The benefit of cardiac resynchronization therapy is not hindered by the number of comorbidities. *J Am Coll Cardiol*. 2017 Oct 17;70(16):2096-2097.

[Ref sup 7] Garcia R, Boveda S, Defaye P, et al. On behalf of the DAI-PP Investigators. Early mortality after implantable cardioverter defibrillator: Incidence and associated factors. *Submitted*

[Ref sup 8] Anantha Narayanan M, Vakil K, Reddy YN, et al. Efficacy of implantable cardioverter-defibrillator therapy in patients with nonischemic cardiomyopathy: a systematic review and meta-analysis of randomized controlled trials. *JACC Clin Electrophysiol* 2017;3:962-970. doi: 10.1016/j.jacep.2017.02.006.

Figure legends

Figure 1. Survival curves for mortality according to early complication occurrence.

Figure 2. Survival curves comparing cardiac resynchronization therapy with a ICD (CRT-D) versus cardiac resynchronization therapy with a pacemaker (CRT-P), using Cox regression with adjustment on the propensity score and all mortality predictors. CI: confidence interval; HR: hazard ratio; N: number.

Figure 3. Survival probability of inappropriate shock occurrence according to implantable cardioverter ICD settings. VH-RATE: very-high-rate group; 1-zone: one therapy zone; 2-zone: two therapy zone.

Figure 4. Proportions of appropriate and inappropriate therapies during follow-up, according to the type of device: single-chamber ICD (SC-ICD) versus dual-chamber ICD (DC-ICD).

Table 1 Main findings.

DAI-PP main findings	<p>Over a decade, age at implantation and the proportions of CRT implantations, NICM and female recipients have increased [6].</p> <p>Compared with males, female recipients of an ICD have a 0.59-fold lower incidence of appropriate ICD therapies [7].</p> <p>Although patients in NYHA class III have a higher overall mortality, driven by an increase in cardiovascular death, they experience a similar incidence of appropriate ICD therapies [8].</p> <p>More than a quarter were implanted without fulfilling proper guideline criteria. Despite a less severe clinical profile at implantation and lower mortality rates, off-guideline patients had, rather surprisingly, similar rates of appropriate therapies [Ref sup 3].</p> <p>After ICD implantation, older patients exhibit higher global mortality, but 19.4% experience at least one appropriate therapy before death, and rates of appropriate therapies are similar overall across age groups [10].</p> <p>Overall, long-term mortality was lower among recipients of CRT with a ICD who had undergone perioperative defibrillation testing (HR 0.6); however, this mainly reflected their less severe cardiac disease [14].</p> <p>Early complication after ICD implantation is associated with a 1.7-fold higher risk of mortality; factors associated with early complications are severe renal impairment, age \geq 75 years, CRT and anticoagulant therapy [11].</p> <p>Programming a very high ventricular tachycardia and fibrillation cut-off rate is associated with a reduction in appropriate therapies and inappropriate shock [22].</p> <p>Compared with single-chamber ICDs, dual-chamber ICDs do not reduce rates of inappropriate therapies, and are associated with a higher rate of complications [23].</p> <p>A risk score including NYHA class \geq III, age > 70 years, QRS duration > 120 ms, atrial fibrillation and GFR < 60 mL/min allows identification of patients at risk of non-arrhythmic death, who might not benefit from ICD therapy [25].</p>
----------------------	--

Electrical storm among patients with an ICD is associated with a 3.77-fold higher risk of death [26].

DAI-PP collaborations
main findings Patients with ICM had better survival when receiving CRT with a ICD compared with those who received CRT alone; no such difference was observed in patients with NICM [19].

A simple clinical risk score called ScREEN, using sex category, renal function, electrocardiogram/QRS width, ejection fraction and NYHA class variables, is able to identify patients who will not respond to CRT [27].

In patients with an indication for CRT, the addition of an ICD conveys additional benefit in well-selected male patients, but possibly not in female patients [20].

The burden of comorbidity did not appear to significantly influence the probability of response to CRT, despite a potentially smaller survival benefit in patients with ≥ 3 comorbidities. Although ICD implantation should be carefully considered in patients with multiple comorbidities, CRT should not be withheld in this context of heart failure patient with enlarged QRS. [Ref sup 6]

Progressive heart failure death was found to represent the most frequent cause of death in patients surviving the first five years after CRT implant. By contrast, SCD represented a very low proportion of late mortality irrespective of the presence of an ICD. [Ref sup 5]

CRT: cardiac resynchronization therapy; GFR: glomerular filtration rate; HR: hazard ratio; ICD: implantable cardioverter defibrillator; ICM: ischaemic cardiomyopathy; NICM: non-ischaemic cardiomyopathy; NYHA: New York Heart Association.

